



UNITED STATES
NUCLEAR REGULATORY COMMISSION
WASHINGTON, D.C. 20555-0001

December 10, 2024

ALL AGREEMENT STATES, CONNECTICUT, AND INDIANA

SUBJECT: NOTIFICATION OF ISSUANCE OF EYE90 MICROSPHERES® LICENSING GUIDANCE (STC-24-065)

Purpose: To inform the Agreement States, Connecticut, and Indiana that the U.S. Nuclear Regulatory Commission (NRC) issued the Eye90 microspheres® Licensing Guidance.

Background: On May 30, 2023, ABK Biomedical, Inc. announced it had received Investigational Device Exemption (IDE) approval from the U.S. Food and Drug Administration (FDA) for Eye90 microspheres®. Eye90 microspheres® are radiopaque yttrium-90 (Y-90) glass microspheres used for treatment of conditions of the liver. Similar to existing microsphere technologies, Eye90 microspheres® are licensed under Title 10 of the Code of Federal Regulations (10 CFR) Part 35, Subpart K, "Other Medical Uses of Byproduct Material or Radiation from Byproduct Material." Like other Y-90 microsphere technologies, Eye90 microspheres® are delivered under fluoroscopy guidance. The radiopacity of the microspheres allows for imaging under fluoroscopy during treatment or after treatment using any X-ray imaging technique. As the use of Eye90 microspheres® requires use of non-licensed sources of radiation, licensees should remain aware of the occupational monitoring requirements for exposure from licensed and non-licensed sources. More information can be found in IN 2021-02 ([ML21152A239](#)).


Enclosure 1 provides the final 10 CFR 35.1000 licensing guidance for Eye90 microspheres® ([ML24247A254](#)). This guidance lists an approved set of regulations and licensing conditions specific to the Eye90 microspheres®. This guidance should be used in concurrence with NUREG-1556, Volume 9, Revision 3, "Consolidated Guidance About Material Licenses: Program-Specific Guidance about Medical Use Licenses," which provides overall licensing guidance for all medical uses of byproduct material. Enclosure 2 provides a supporting technical analysis table which includes a list of 10 CFR Part 35 regulations and conditions the NRC has deemed acceptable for the use of Eye90 microspheres®. However, as stated in the licensing guidance, applicants may submit alternative list of regulations and specific conditions to be reviewed on a case-by-case basis.

The attached licensing guidance will be posted on NRC's public Web site (<https://www.nrc.gov/materials/miau/med-use-toolkit/emerg-licensed-med-tech.html>).

POINT OF CONTACT: Sarah Spence E-MAIL: sarah.spence@nrc.gov
TELEPHONE: (301) 415-3697

If you have any questions regarding this correspondence, please contact me at (301) 415-3521 or the individual named below:

Sincerely,



Signed by Giantelli, Adelaide
on 12/10/24

Adelaide Giantelli, Chief
State Agreement Liaison Programs Branch
Division of Materials Safety, Security, State,
and Tribal Programs
Office of Nuclear Material Safety
and Safeguards

Enclosures:

1. Eye90 microspheres® Licensing Guidance
2. Consolidated Technical Analysis

STC 24 065 Notification Eye90 microspheres Licensing Guidance DATE December 12, 2024

DISTRIBUTION:

ADAMS Accession No.: ML24325A436; Ltr ML24325A439

OFFICE	NMSS/MSST/MSEB	NMSS/MSST /MSEB	NMSS/MSST/MSEB	NMSS/MSST/SMPB
NAME	SSpence	SSKTapp	KTCEinberg	CEAGiantelli
DATE	Dec 2, 2024	Dec 4, 2024	Dec 5, 2024	Dec 10, 2024

OFFICIAL RECORD COPY

Yttrium-90 Microsphere Brachytherapy Sources and Devices Eye90 Microspheres® Licensing Guidance

September 27, 2024

NRC Contact

Sarah Spence

MedicalQuestions.Resource@nrc.gov

Table of Contents

1. Background..... 1

2. 10 CFR 35.1000 Use..... 1

3. Licensing Guidance..... 2

4. General 2

 4.1 Requirements not Specific to 10 CFR 35.1000 Use 2

 4.2 Radionuclides, Form, Possession Limits, and Purpose of Use..... 3

 4.3 Facility Address and Description 3

 4.4 Leak Tests 3

5. Training and Experience..... 4

 5.1 Authorized Users 4

 5.2 Radiation Safety Officer 7

 5.3 Training and Experience Documentation 8

 5.4 Team Approach 8

 5.5 Notification..... 9

6. License Commitments..... 9

 6.1 Procedures for Administration 10

 6.2 Written Directives..... 10

 6.3 Medical Event Reporting 12

 6.4 Sealed Source and Device Use 13

 6.5 Inventory..... 13

 6.6 Labeling 14

 6.7 Patient Release 14

 6.8 Surveys..... 14

 6.9 Radiation Protection Program Changes 14

7. Notes to Licensees..... 15

7.1	Change in Physical Conditions of Use	15
7.2	Use of Other Y-90 Microspheres.....	16
7.3	Waste Disposal Issues.....	16
7.4	Explanted Tissues, Autopsy and Cremation.....	17
7.5	Radiation Safety Committee	18
8.	Notes to Regulators.....	18
8.1	Inspection Frequency.....	18
8.2	Program Code	18
9.	Paperwork Reduction Act Statement.....	18
10.	Public Protection Notification	19

1. Background

This licensing guidance for yttrium-90 (Y-90) microsphere brachytherapy is exclusively for the use of Eye90 microspheres[®]. At the time of publication of this guidance, Eye90 microspheres[®] are a new microsphere device approved by the Food and Drug Administration (FDA) under an Investigational Device Exemption (IDE) for clinical trial.

Eye90 microspheres[®] are permanent implantable radiopaque glass Y-90 microspheres used for treatment of conditions of the liver. The microspheres are provided as a unit dose in 2.5 mL saline, contained in an acrylic shield. Following angiographic pre-therapy evaluation, the microspheres are delivered to tumor vasculature via a catheter placed under fluoroscopy guidance using a proprietary delivery system comprised of a sterile, single-use delivery device and a re-useable, nonsterile system container. This product is unique to the existing FDA-approved Y-90 microsphere products at the time of publication, due to the radiopaque quality of the microspheres. The Authorized User (AU) may choose to use fluoroscopy to directly image the radiopaque microspheres in the tumor vasculature during administration and via any x-ray imaging modality thereafter.

This licensing guidance has been adapted from the existing guidance for other microsphere brachytherapy, “Yttrium-90 Microsphere Brachytherapy Sources and Devices TheraSphere[®] and SIR-Spheres[®] Licensing Guidance” (ML20080J208), to reflect changes unique to Eye90 microspheres[®], and, in particular, to reflect the ability to visually observe the microspheres during administration via fluoroscopy. Because this product is new and still being investigated in clinical trials at the time of publication, this licensing guidance was needed to facilitate any necessary future revisions as the NRC and medical communities gain experience with Eye90 microspheres[®].

2. 10 CFR 35.1000 Use

Although Eye90 microspheres[®] are manual brachytherapy sources used for permanent implantation therapy, Eye90 microspheres[®] have many unique properties that merit radiation safety considerations other than those required by Title 10 of the *Code of Federal Regulations* (10 CFR) Part 35, “Medical Use of Byproduct Material,” Subpart F, “Manual Brachytherapy.” These unique properties include the microspheres’ small size, the large number of microspheres used in a treatment, and the route of administration. As a result, Eye90

microspheres[®] brachytherapy is regulated under 10 CFR 35.1000, “Other medical uses of byproduct material or radiation from byproduct material¹.”

3. Licensing Guidance

This guidance provides applicants with an acceptable means of satisfying the requirements for a license for the use of Eye90 microspheres[®] and is not intended to be the only means of satisfying the requirements for a license. The applicant must submit the information required to meet 10 CFR 30.33 and 35.12, as described below. The applicant should submit additional information and commitments requested below or may, unless the information is specifically required by regulation, submit alternative information and commitments for review by the NRC to make a licensing determination. The commitments incorporated into the license by license condition will be reviewed during routine inspections. If an applicant commits to the guidance provided below, the applicant is committing to follow commitments described with the use of the word “should.”

4. General

4.1 Requirements not Specific to 10 CFR 35.1000 Use

Applicants must commit to meet the general requirements in 10 CFR Part 35, Subpart A, “General Information;” Subpart B, “General Administrative Requirements;” Subpart C, “General Technical Requirements;” Subpart L, “Records;” Subpart M, “Reports;” and Subpart N, “Enforcement;” except as specified in this guidance. Additionally, applicants must meet applicable requirements of 10 CFR Part 19, “Notices, Instructions and Reports to Workers: Inspection and Investigations;” Part 20, “Standards for Protection Against Radiation;”² Part 30, “Rules of General Applicability to Domestic Licensing of Byproduct Material;” and Part 71, “Packaging and Transformation of Radioactive Material.”

¹ 10 CFR 35.1000 is designated as Compatibility Category D. Agreement States are not required to adopt these regulations for purposes of compatibility but are not prohibited from adopting Compatibility Category D regulations if they so choose. If Agreement States choose to adopt this licensing guidance, references to 10 CFR should be changed to the equivalent Agreement State regulations.

² Refer to [IN-21-02](#) for further information regarding compliance with Part 20. Note that administration of Eye90 microspheres[®] may result in more fluoroscopy time in a given procedure compared to similar microsphere technologies.

Applicants may refer to NUREG-1556, Volume 9, Revision 3, “Consolidated Guidance About Material Licenses: Program-Specific Guidance About Medical Use Licenses,” as it provides overall licensing guidance for all medical uses of byproduct material, including information on how to submit a facility’s address and description and applicable model procedures for audits, occupational dose monitoring programs, and surveys.

4.2 Radionuclides, Form, Possession Limits, and Purpose of Use

Pursuant to 10 CFR 35.12, the applicant shall identify the radionuclide, chemical/physical form, requested maximum possession limit, and purpose of use. This information may be submitted under a signed, dated letter or NRC Form 313, “Application for Materials License.” The following table provides the format for an acceptable request.

	Eye90 Microspheres®
Radionuclides (NRC Form 313 Item 5a)	Yttrium-90
Chemical/Physical Form (NRC Form 313 Item 5b)	Glass microsphere (current manufacturer as listed in the Sealed Source and Device Registry [e.g., ABK Biomedical, Inc. Model Eye90 microspheres®])
Maximum Possession Limit (NRC Form 313 Item 5c)	X* Ci total
Purpose of Use (NRC Form 313 Item 6)	Eye90 microspheres® for permanent brachytherapy using delivery system as listed in the Sealed Source and Device Registry

* Based on the maximum amount the applicant anticipates having at one time (i.e., 3 Ci)

4.3 Facility Address and Description

Provide an address of use and description of the location where the Eye90 microspheres® will be used and stored.

4.4 Leak Tests

Leak tests are not required for Eye90 microspheres[®]. The small size and large number of Y-90 microspheres make leak testing, as required by [10 CFR 35.67\(b\)](#), impractical. Further, leak testing is not required as the activity of each Y-90 microsphere is below the threshold in [10 CFR 35.67\(f\)\(3\)](#).

5. Training and Experience

5.1 Authorized Users

NRC has determined that individuals meeting the AU training and experience (T&E) criteria A, B, or C provided below can be authorized for the use of Eye90 microspheres[®] brachytherapy. Applicants may also submit alternative T&E criteria to be reviewed on a case-by-case basis by NRC staff. The alternative T&E commitments should include an explanation of why the applicant believes the alternative T&E commitments demonstrate that the individuals are qualified to be an AU.

A. Has obtained proper training and experience, i.e.:

1. Is identified as an AU for medical use in [10 CFR 35.1000](#) for Y-90 microspheres, [10 CFR 35.400](#), "Use of sources for manual brachytherapy," or for medical uses in [10 CFR 35.300](#), "Use of unsealed byproduct material for which a written directive is required," that includes the use described in 10 CFR 35.390(b)(1)(ii)(G)(3) on one of the following licenses or permits that authorizes the medical use of byproduct material: A Commission or Agreement State license, a permit issued by a Commission master materials licensee, a permit issued by a Commission or Agreement State licensee of broad scope, or a permit issued by a Commission master materials license broad scope permittee; or
2. Meets the training and experience requirements of [10 CFR 35.390](#) or [10 CFR 35.490](#); or
3. Meets the training and experience guidelines as follows:
 - i. Experience in diagnostic radiology demonstrated by:
 - (a) Board certification in interventional radiology/diagnostic radiology by the American Board of Radiology (ABR)³; or
 - (b) Board certification in diagnostic radiology by the ABR³[Error! Bookmark not defined.](#); or

³ For the purposes of criteria A.3.i.a. and A.3.i.b., these certificates do not need to be recognized by the NRC as listed on NRC's [Medical Uses Licensee Toolkit Web page](#) (i.e., "AU eligible").

- (c) Board certification in diagnostic radiology by the American Osteopathic Board of Radiology (AOBR)³; or
 - (d) Three years supervised clinical experience in diagnostic radiology; and
 - b. Experience in interventional radiology demonstrated by:
 - (a) Board certification in interventional radiology/diagnostic radiology by the ABR³; or
 - (b) Board subspecialty certification in interventional radiology by the AOBR³; or
 - (c) One year of supervised clinical experience in interventional radiology; and
- ii. Has 80 hours of classroom and laboratory training⁴ for byproduct material requiring a written directive, applicable to Y-90 microspheres, which may be concurrent with training received in accordance with criterion A.3.i in:
 - a. Radiation physics and instrumentation; and
 - b. Radiation protection; and
 - c. Mathematics pertaining to the use and measurement of radioactivity; and
 - d. Radiation biology; and
- iii. Has work experience under the supervision of an AU for Eye90 microspheres[®] brachytherapy or training provided by an Eye90 microspheres[®] manufacturer representative involving:
 - a. Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys; and
 - b. Performing quality control procedures on instruments used to determine the activity of Eye90 microspheres[®] and performing checks for proper operation of survey meters; and
 - c. Calculating and measuring the activity and safely preparing the Eye90 microspheres[®] to be delivered to the patient or human research subject; and
 - d. Using procedures to control and to contain spilled byproduct material, including Eye90 microspheres[®], safely and using proper decontamination procedures. The procedures should address any special circumstances that may be encountered, such as the electrostatic charge of Eye90 microspheres[®] and the proper survey instrument and survey technique for beta emitters⁵; and

⁴ For Board Certified physicians, if the Board Certification is recognized by the NRC on the NRC's Medical Uses Licensee Toolkit Web page for 10 CFR 35.290, 35.390, 35.392, 35.394, and 35.396, the applicant or licensee need not submit detailed documentation of those AUs' classroom and laboratory training to satisfy section A.3.ii. The applicant or licensee need only confirm that the individual has completed training on the use of Y-90 microspheres.

⁵ [Appendix N, "Model Emergency Procedures," NUREG-1556, Volume 9](#) provides additional guidance.

- iv. Has work experience or training under the supervision of an AU or manufacturer representative⁶ for Eye90 microspheres[®] brachytherapy, including:
 - a. Preparing and administering patient dosage. The individual does not have to be the physician who places the micro-catheter or administers patient dosage, but it is necessary that the individual have training in the administration process, including selection of activity of Eye90 microspheres[®] to be administered to each treatment site and catheter positioning to ensure administration of the Eye90 microspheres[®] is in accordance with the written directive; and
 - b. Using administrative controls to prevent a medical event involving the use of byproduct material⁷; and
 - c. Evaluation of patient or research subject's treatments to determine whether the administered dosage was in accordance with the written directive or if a medical event has occurred.

B. Has successfully completed training in the operation of the delivery system, safety procedures, and clinical use for Eye90 microspheres[®]. This requirement may be satisfied by completing a training program provided by the vendor for new users or by receiving training supervised by an AU⁸ who is authorized for Eye90 microspheres[®]. Clinical use training to support unsupervised use must include at least three hands-on patient cases for Eye90 microspheres[®], conducted in the physical presence of an AU^{6,8} who is authorized for Eye90 microspheres[®].

Additionally, if the proposed AU is already trained on the use of another model of microspheres, they must be trained in the specific differences between Eye90 microspheres[®] and the other model(s) of microspheres for which they are approved, including activity prescription, written directive preparation, microsphere administration, and spill procedures.

However, if a proposed AU cannot complete patient cases prior to authorization; the licensee may request conditional approval with the proposed AU's completion of at least three mock

⁶ As Eye90 Microspheres[®] are in clinical trials and there are minimal licensees approved at the time this licensing guidance was initially published in 2024, there are a limited number of AUs available to provide training and available to sign attestations. Therefore, training and attestations provided by a manufacturer representative will be accepted in lieu of by an AU until September 27, 2026.

⁷ Appendix S, "Model Procedures for Developing, Maintaining, and Implementing Written Directives," NUREG-1556, Volume 9 provides additional guidance.

⁸ A physician who is not an AU but who meets the T&E criteria for Eye90 microspheres may be approved to supervise another physician's training or first hands-on patient cases on a case-by-case basis.

simulated cases. Mock simulated cases should demonstrate issues that are encountered during Eye90 microspheres[®] administration procedures and should be completed by the individual in the physical presence of a manufacturer representative or an AU⁶ who is authorized Eye90 microspheres[®]. Following conditional approval, the individual should complete the clinical casework described at the beginning of this section within a year following the license issuance or amendment that names the individual as an AU for Eye90 microspheres[®] use. The licensee may submit documentation to the NRC requesting an extension of this timeframe. The supporting documentation should include a commitment to perform continuing T&E (e.g., one additional mock case prior to performing patient cases) in Eye90 microspheres[®] until the first three patient cases are completed, and

- C. Has obtained written attestation that the individual has satisfactorily completed the requirements in criteria A and B of this section and is able to independently fulfill the radiation safety-related duties as an AU for Eye90 microspheres[®]. The attestation must be obtained from either:
1. An AU^{6,8} who is authorized for Eye90 microspheres[®]; or
 2. A residency or fellowship program director who affirms in writing that the attestation represents the consensus of the residency or fellowship program faculty where at least one faculty member is a physician who is an AU⁶ for Eye90 microspheres[®] brachytherapy and concurs with the attestation provided by the residency or fellowship program director. The residency or fellowship training program must be approved by the Residency Review Committee of the Accreditation Council for Graduate Medical Education or the Royal College of Physicians and Surgeons of Canada or the Council on Postdoctoral Training of the American Osteopathic Association and must include T&E specified in criteria A and B of this section.

In accordance with 10 CFR 35.59, the T&E specified above must have been obtained within the seven years preceding the date of application or the individual must have had related continuing education and experience since the required T&E was completed. Recent training provided under Section B may be sufficient to show recentness of training. This recentness of training requirement applies to all individuals, including those who are board certified or listed as an AU on an NRC or Agreement State license.

5.2 Radiation Safety Officer

The Radiation Safety Officer (RSO) must have training as specified in 10 CFR 35.50, including training in radiation safety, regulatory issues, and emergency procedures for Y-90 microsphere use. An RSO already listed on a license that includes one type of Y-90 microsphere device does not require additional approval for another type of Y-90 microsphere device but should be familiar with all radiation safety aspects, including cleaning up spills, associated with all devices used at the facility.

5.3 Training and Experience Documentation

The applicant must submit documentation of the above T&E for all physicians requesting authorization to use Eye90 microspheres®. This documentation shall include the clinical use cases and written attestation and supervising physician T&E, if necessary. For individuals completing the patient cases following the license amendment, this documentation shall include documentation from the manufacturer representative or supervising physician of the three mock simulated cases and a commitment that each individual will complete at least the first three hands-on patient cases supervised in the physical presence of an AU^{6,8} who is authorized for the Eye90 microspheres®. The documentation should commit to requiring the individual to perform a mock simulated case if six months or more have passed since last performing a hands-on or mock simulated case. Additionally, for applicants that have individuals completing the patient cases following the license amendment, the applicant's commitment will include submitting documentation from the manufacturer to the appropriate NRC Regional Office within 60 days of when these three patient cases have been satisfactorily completed.

5.4 Team Approach

Microsphere brachytherapy treatment is usually conducted using a multi-disciplinary team approach. The AU should consult with individuals, as necessary, with expertise in:

- cancer management (e.g., radiation or medical oncology);
- catheter placement;
- radiation dosimetry; and
- safe handling of unsealed byproduct material.

One individual may satisfy more than one of the listed areas of expertise. The applicant shall commit to provide training in the licensee's procedures to all individuals involved in Eye90

microspheres[®] use, commensurate with the individual's duties to be performed. This training must be provided to all individuals ordering, preparing, measuring, performing dosimetry calculations, or administering Eye90 microspheres[®].

5.5 Notification

The NRC recognizes that, if an AU satisfies the T&E listed in NRC's licensing guidance for Eye90 microspheres[®] and is currently listed on a Commission or Agreement State medical use license or permit for Eye90 microspheres[®], the AU should be allowed to work under a different license for the medical use of Eye90 microspheres[®]. A limited specific medical use applicant initially applying for authorization for the medical use of Eye90 microspheres[®] or an existing licensee applying for an amendment may request authorization to notify the NRC in the future of AUs that it has permitted to work at its facility, without requesting an additional license amendment, provided the following conditions are met:

1. the AU satisfies the T&E listed in this licensing guidance for Eye90microspheres[®]; and
2. the AU is currently listed for Eye90 microspheres[®] use on a Commission or Agreement State license, a permit issued by a Commission master materials licensee, a permit issued by a Commission or Agreement State licensee of a broad scope, or a permit issued by a Commission master materials license broad scope permittee; and
3. the licensee provides the NRC a copy of the license or permit on which the AU is listed for Eye90 microspheres[®] use; and
4. the licensee provides the NRC documentation of the completion of three patient cases if previously not submitted to the NRC; and
5. the licensee provides documentation of the above listed conditions to NRC for each AU no later than 30 days after the date that the licensee allows the AU to work as an AU for the Eye90 microspheres[®].

If this authorization is approved, these notification conditions will be incorporated as license conditions on the license.

6. License Commitments

The applicant shall commit to follow all the requirements in 10 CFR Part 35 for brachytherapy sources and manual brachytherapy use, except where replaced by the following licensing commitments:

6.1 Procedures for Administration

The licensee must have procedures for administration requiring a written directive as specified in [10 CFR 35.41](#), specifically to ensure high confidence that the patient's or human research subject's identity is verified before each administration and each administration is in accordance with the written directive. As Eye90 microspheres[®] are too small to be calibrated in accordance with [10 CFR 35.432](#), the licensee shall determine and record the activity of each dosage before medical use in accordance with [10 CFR 35.63](#) and [10 CFR 35.60](#) even though Eye90 microspheres[®] are not considered to be unsealed byproduct material. The licensee shall commit to following the manufacturer's procedures or submit alternative methods for calculating and documenting the dose or activity to the treatment site; preparing the dose for administration; determining shunting to non-treatment sites; and determining if a medical event has occurred (e.g., performing pre- and post-vial dose measurements with appropriate instrumentation, evaluating post-treatment imaging). For the purpose of this guidance, shunting is defined as blood flow through pathway or bypass due to patient vasculature causing the Eye90 microspheres[®] to flow to an unwanted location. Unexpected dose or activity to an organ or tissue other than the treatment site that is caused by catheter placement during delivery of the Eye90 microspheres[®] is not considered shunting and should be evaluated as a possible medical event.

Administration of Eye90 microspheres[®] must be performed in accordance with the written directive. The licensee shall record the absorbed dose or activity delivered to the treatment site and should record the absorbed dose to non-target sites as appropriate. As described in Section 6.2 Written Directives, the choice of using activity or dose should be consistent across all documentation and evaluations. The record shall be prepared within 24 hours after the completion or termination of the administration and must include the name of the individual who determined the dose or administered activity and the date the record is completed.

6.2 Written Directives

The licensee must complete a written directive, which must be dated and signed by an AU before the administration in accordance with [10 CFR 35.40\(a\)](#) and [10 CFR 35.40\(c\)](#) unless a delay in order to provide a written directive would jeopardize the patient's health, as allowed under [10 CFR 35.40\(c\)\(1\)](#). The licensee shall retain a copy of the written directive in accordance with [10 CFR 35.2040](#).

Due to the unique properties of Eye90 microspheres[®] brachytherapy, the following written directive condition should be used instead of [10 CFR 35.40\(b\)](#).

The written directive shall include the patient or human research subject's name; the treatment site; the radionuclide (including the physical form [Y-90 microspheres]); the model of spheres (e.g. Eye90 microspheres[®]); the prescribed dose or activity; and, if appropriate for the type of microsphere used, the statement "or dose or activity delivered at stasis."

For the purpose of written directive and medical event reporting requirements in the Eye90 microspheres[®] guidance, prescribed activity (mCi or GBq) means the total activity intended to be administered whereas "prescribed dose" means the total planned dose (rad or Gy). The choice of prescribed activity or prescribed dose should be used consistently for all subsequent documentation and evaluations. As described in 10 CFR 35.2, "treatment site" means the anatomical description of the tissue intended to receive a radiation dose, as described in a written directive. For instance, the treatment site may be described as the lobe or segment that is intended to receive the Eye90 microspheres[®] and the tissue that is expected to receive Eye90 microspheres[®] due to shunting. For the purpose of this guidance, stasis is defined as a stoppage or slowdown in the flow of blood. The inability to complete administration due to clogging or kinking of the catheter is not considered stasis.

6.2.1 Termination of Treatment Due to Stasis

If the administration was terminated because of stasis, then the total dose or activity to the treatment site is the value of the total dose or activity administered when stasis occurred, and the administration was terminated. The record shall be prepared within 24 hours after the completion or termination of the administration and must include the name of the individual who

determined the administered dose or activity, the signature of an AU for Eye90 microspheres[®], and the date signed.

6.2.2 Emergent Patient Conditions

If the procedure must be modified due to emergent patient conditions that prevent administration in accordance with the written directive (e.g., artery spasm or sudden change in blood pressure), the AU shall document such changes in the written directive within 24 hours after the completion or termination of the administration. The modification to the written directive shall include the reason for not administering the intended dose or activity, the signature of an AU for Eye90 microspheres[®], and the date signed.

6.2.3 Termination of Treatment Due to Observed Deposition

If the AU decides to terminate administration due to observation of undesired radiopaque microsphere deposition, the AU may make an oral revision to a written directive to modify the prescribed dose or activity. The oral revision must be documented as soon as possible in the patient's record. A revised written directive must be signed by an AU within 24 hours of the oral revision and shall include a reason for the termination, the dose or activity delivered, the signature of an AU for Eye90 microspheres[®], and the date signed. If the dose delivered to the unintended target exceeds the thresholds for medical event reporting as outlined in Section 6.3, then the licensee shall comply with the medical event reporting and notification requirements as described in [10 CFR 35.3045\(b\)-\(g\)](#).

6.3 Medical Event Reporting

In place of 10 CFR 35.3045(a), the licensee shall commit to report any event, except for an event that is caused by shunting as described in the criteria below, or as a result of patient intervention, as defined in 10 CFR 35.2 as an action by the patient or human research subject, whether intentional or unintentional, such as dislodging or removing treatment devices or prematurely terminating the administration. The criteria for event reporting is:

- the administration of byproduct material results in a dose that exceeds 0.05 Sv (5 rem) effective dose equivalent or 0.5 Sv (50 rem) dose equivalent to an organ or tissue per 10 CFR 35.3045; and

- an administration of the wrong radionuclide or type of microsphere; or
- an administration to the wrong individual or human research subject; or
- an administration by the wrong route of administration; or
- an administration by the wrong mode of treatment; or
- the total dose or activity delivered differs from the prescribed dose or activity, as documented in the written directive, by 20 percent or more, except when stasis or emergent patient conditions are documented and resulted in a total dose or activity administered that was less than that prescribed; or
- A dose equivalent to the skin or an organ or tissue other than the treatment site that exceeds by 0.5 Sv (50 rem) to an organ or tissue and 50 percent or more of the dose expected from the administration defined in the written directive (excluding shunting as defined in Section 6.1 when shunting was evaluated prior to the treatment in accordance with the manufacturer's procedures.)

Additionally, the licensee shall comply with the medical event reporting and notification requirements as described in [10 CFR 35.3045\(b\)-\(g\)](#).

6.4 Sealed Source and Device Use

The licensee should commit to only use Eye90 microspheres[®] for therapeutic medical uses as approved in the Sealed Source and Device (SSD) registration certificate for Eye90 microspheres[®], including maximum activity per vial limit.

6.5 Inventory

Due to the short half-life of Y-90 (64 hours) and the fact that microspheres are not managed as individual discrete sources, the requirements in [10 CFR 35.67](#) for semi-annual physical inventory of brachytherapy sources and recordkeeping in [10 CFR 35.2406](#) are not applicable to microspheres. Rather, the requirements for brachytherapy source accountability ([10 CFR 35.406](#)), receipt ([10 CFR 20.1906](#)), labeling ([10 CFR 20.1904](#) and [10 CFR 35.69](#)), storage ([10 CFR 20.1801](#) and [10 CFR 35.92](#)), and disposal (see the “Waste Disposal Issues” section of this guidance document) are sufficient to ensure accountability of Y-90 in the form of microspheres possessed by a licensee.

6.6 Labeling

The licensee should commit to the following when the Eye90 microspheres[®] are placed in vials, syringes, or radiation shields that are not labeled by the manufacturer:

- Label vials and vial radiation shields with the radioactive device (i.e. Eye90 microspheres[®]); and
- Label syringes and syringe radiation shields with the radioactive device.

6.7 Patient Release

The licensee should commit to develop procedures that describe measures taken to ensure that radiation emissions, which may include bremsstrahlung, from each patient or human research subject permits his or her release in accordance with [10 CFR 35.75](#). Guidance for release of patients or human research subjects following administration of radioactive materials may be found in [Regulatory Guide 8.39](#), "Release of Patients Administered Radioactive Materials."

6.8 Surveys

As the Eye90 microspheres[®] are too small to be seen, licensees should survey, with an appropriate calibrated radiation detection survey instrument (per 10 CFR 35.61), all areas that the Eye90 microspheres[®] are prepared for use or administered. The survey should be conducted immediately following each preparation and administration in unrestricted areas and by the end of the day for restricted areas. A licensee should retain a record of each survey for three years and the record should include the date of the survey, the results of the survey, the instrument used to perform the survey, and the name of the individual who performed the survey. Licensees do not need to perform surveys in an area(s) where patients or human research subjects are confined when they cannot be released under [10 CFR 35.75](#).

6.9 Radiation Protection Program Changes

This guidance may be revised as additional experience is gained regarding the medical use of Eye90 microspheres[®]. A licensee currently authorized to use these products that is committed by license condition to following provisions in a previous revision of this guidance may request a license amendment to commit to following this revision of the guidance instead. The licensee

must apply for and receive this license amendment in order to make program changes to conform to this revision of the guidance.

An applicant initially applying for authorization for the medical use of Eye90 microspheres[®], or a licensee applying for an amendment to conform with this revision of the guidance may request to incorporate into its license a change process similar to [10 CFR 35.26](#). Such a change process can allow some future changes to radiation safety programs without a license amendment provided that the change process requires the following conditions to be met for revisions to the radiation safety program:

1. the revision is in compliance with the regulations; and
2. the revision is based upon NRC's current guidance for Eye90 microspheres[®] Y-90 microspheres 35.1000 use posted on the [NRC's Medical Uses Licensee Toolkit Web page](#); and
3. the revision has been reviewed and approved by the licensee's RSO and licensee's management; and
4. the affected individuals are instructed on the revised program before the change is implemented; and
5. the licensee will retain a record of each change for five years; and
6. the record will include a copy of the appropriate website guidance, the old procedure, the new procedure, the effective date of the change, and the signature of the licensee management that reviewed and approved the change.

If approved, these conditions for use of the updated guidance will be incorporated as license conditions in the license.

7. Notes to Licensees

7.1 Change in Physical Conditions of Use

If the physical conditions of use exceed those reported in the SSD registration certificate, the limited specific medical use licensee shall request an amendment for the new conditions, and a broad scope licensee shall perform its own engineering and radiation safety evaluation addressing those differences.

7.2 Use of Other Y-90 Microspheres

The SSD safety evaluation for a specific manufacturer's Y-90 microsphere does not cover the use of any other Y-90 microspheres in its device for administration. The medical use of such a source will require a new SSD registration certificate (or safety evaluation by the broad scope medical use licensee) that addresses the conditions of use, safety of the new Y-90 microspheres, and compatibility of the new microspheres with the microsphere delivery system(s).

The SSD safety evaluation for a given manufacturer's Y-90 microsphere delivery system does not cover the use of that manufacturer's Y-90 microspheres with another manufacturer's delivery system or the use of another manufacturer's Y-90 microspheres with the given manufacturer's delivery system. Before authorization, the medical use of such a delivery system will require a new SSD registration certificate (or safety evaluation by the broad scope medical use licensee) that addresses the conditions of use, safety of the microsphere delivery system, and compatibility of the new delivery system with the Y-90 microspheres.

7.3 Waste Disposal Issues

Y-90 microspheres are known to potentially contain radioactive impurities, some of which are long-lived (i.e., half-lives of greater than 120 days) (Refer to [Information Notice \(IN\) 2007-10](#), "Yttrium-90 Therasphere® and Sirspheres® Impurities"). Due to different manufacturing processes, the activity and radionuclides of the impurities vary for different Y-90 microsphere products. Impurities that have been recently found in reactor-activated microspheres include small amounts of long-lived radionuclides such as europium-152, europium-154, and cobalt-60.⁹ Impurities that have been recently found from microspheres with generator-produced Y-90 include trace amounts of strontium-90.¹⁰ While this IN was generated in response to other

⁹ J. Metyko, J. Williford, W. Erwin, J. Poston, S. Jimenez. "Long-lived Impurities of ⁹⁰Y-labeled microspheres, TheraSphere and SIR-Spheres, and the impact on patient dose and waste management." *Health Phys.* **103**(3), S204-S208 (2012).

¹⁰ J. Metyko, W. Erwin, J. Poston, and S. Jimenez. "⁹⁰Sr Content in ⁹⁰Y-labeled SIR-Spheres and Zevalin." *Health Phys.* **107**(5), S177-S180 (2014)

products, licensees should be prepared to follow similar precautions and procedures when handling waste generated by Eye90 Microspheres®.

Licensees should be aware that the activity and type of impurities can change and be different from that described above. The NRC does not limit manufacturers to specific manufacturing processes, and it is therefore possible for the activity and types of radionuclide impurities to change for all Y-90 microsphere products. Additionally, unused or partially used vials are likely to contain higher activities of impurities.

Although impurities need not be listed on an NRC license, licensees are responsible to ensure the microspheres are handled and disposed of in accordance with 10 CFR [Part 20](#) and [Part 35](#) requirements. Specifically, [10 CFR 35.92](#) requires that licensees monitor byproduct material with a physical half-life of less than or equal to 120 days at the surface before disposal and determine that its radioactivity cannot be distinguished from the background radiation level with an appropriate radiation detection survey meter before disposal. Therefore, regardless of the length of time they have been allowed to decay, licensees are not permitted to dispose of Y-90 microspheres if radioactivity can be distinguished from the background radiation level with an appropriate radiation detection survey meter.

If waste is determined to contain impurities with a physical half-life of greater than 120 days that can be distinguished from the background radiation level with an appropriate radiation detection survey meter, the licensee may need to use one or more of the following means to dispose of waste associated with the Eye90 microspheres®:

- return the Eye90 microspheres® to the manufacturer, if the manufacturer is authorized to receive Y-90 microspheres; or
- transfer the Eye90 microspheres® to an authorized recipient pursuant to requirements in 10 CFR [Part 20](#) and [Part 30](#).

See [Regulatory Issue Summary 2004-17, Revision 1](#), “Revised Decay-in-Storage Provisions for the Storage of Radioactive Waste Containing Byproduct Material,” for more information regarding requirements for holding waste for decay-in-storage.

7.4 Explanted Tissues, Autopsy and Cremation

Eye90 microspheres® are permanent implants that are not removed from the body by biological methods. Because Y-90 has a 64-hour half-life, Y-90 will likely have significantly decayed before a patient's death. Patients treated with Eye90 microspheres® will not usually represent an external radiation hazard to persons handling the body. However, when managing explanted tissues treated with Eye90 microspheres®, or in the case of autopsy or cremation, a radiation hazard exists for individuals who handle tissues that may contain radioactive material, especially if the event of explantation or death occurs soon after treatment with Eye90 microspheres®. The National Council on Radiation Protection and Measurements (NCRP) [Report No. 155](#), "Management of Radionuclide Therapy Patients," December 2006, may contain helpful information for radiation safety considerations associated with autopsy or cremation of patients with permanent implants. Additionally, [NUREG-1556, Volume 9](#), Appendix N, "Model Emergency Procedures," contains additional guidance regarding autopsy and cremation of patients who have received therapeutic amounts of radionuclides.

7.5 Radiation Safety Committee

If a licensee is required to have an RSC in accordance with 10 CFR 35.24(f), then the committee must include an AU for the use of Y-90 microspheres.

8. Notes to Regulators

8.1 Inspection Frequency

Licenses authorizing Eye90 microspheres® brachytherapy should be inspected every two years. Licenses authorizing emerging technology under 10 CFR 35.1000 are assigned a Priority 2 inspection code, as is explained in Enclosure 1 to [Inspection Manual Chapter 2800](#). The list of the program codes along with the assigned priority codes can be found at <https://www.nrc.gov/materials/miau/mat-toolkits.html>.

8.2 Program Code

The NRC regions should use program code 02240.

9. Paperwork Reduction Act Statement

The information collections contained in this guidance are covered by the requirements of 10 CFR Parts 30 and 35, which were approved by the Office of Management and Budget (OMB), approval numbers 3150-0017 and 3150-0010. Send comments regarding this information collection to the Information Services Branch (T6-A10M), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by e-mail to Infocollects.Resource@nrc.gov, and to the OMB reviewer at: OMB Office of Information and Regulatory Affairs (3150-0017, 3150-0010), Attn: Desk Officer for the Nuclear Regulatory Commission, 725 17th Street, NW Washington, DC 20503; e-mail: oira_submission@omb.eop.gov.

10. Public Protection Notification

The NRC may not conduct or sponsor, and a person is not required to respond to, a request for information or an information collection requirement, unless the requesting document displays a currently valid OMB control number.

Appendix

Consolidated Technical Analysis

The following table provides a list of 10 CFR Part 35 regulations and conditions the NRC has determined are applicable for use of *Eye90 microspheres*[®]. Licensees shall comply with all regulations which address use of *Eye90 microspheres*[®]. The table also provides specific conditions which the NRC has determined are necessary for the medical use of *Eye90 microspheres*[®]. In addition, the table lists where licensees and applicants can find additional guidance. Applicants may submit alternative list of regulations and specific conditions to be reviewed on a case-by-case basis by NRC staff.

Section	Description	Use Addressed in Regulation	Guidance	<i>Eye90 microspheres</i> [®] Guidance Section	Comment
35.1	Purpose and scope	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.2	Definitions	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.5	Maintenance of records	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.6	Provisions for the protection of human research subjects	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.7	FDA, other Federal, and State requirements	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		

Appendix

Consolidated Technical Analysis

Section	Description	Use Addressed in Regulation	Guidance	<i>Eye90 microspheres</i>[®] Guidance Section	Comment
35.8	Information collection requirements: OMB approval	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		

Appendix
Consolidated Technical Analysis

35.10	Implementation	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.11	License required	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.12	Application for license, amendment, or renewal	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.13	License amendments	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.14	Notifications	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.15	Exemptions regarding Type A specific licenses of broad scope	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.18	License issuance	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.19	Specific exemptions	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		

Appendix
Consolidated Technical Analysis

Subpart B – General Administrative Requirements					
35.24	Authority and responsibilities for the radiation protection program	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.26	Radiation protection program changes	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.27	Supervision	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.40	Written directives (WDs)	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.2	35.40(b)(6) details written directive requirements for permanent implant brachytherapy. 35.40(b)(6)(ii) requires a record of the number of sources implanted, which is not possible due to the small size and large number of microspheres involved in Eye90 therapy. Therefore there are different items required by the written directive.
35.41	Procedures for administrations requiring a WD	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.2	The guidance document provides additional information specific to Eye90 Microspheres.
35.49	Suppliers for sealed sources or devices for medical use	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.50	Training for Radiation Safety Officer (RSO) and Associate RSO	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®]	5.2	The guidance document provides additional information specific to Eye90 Microspheres.

Appendix
Consolidated Technical Analysis

			<input type="checkbox"/> Other		
35.51	Training for an authorized medical physicist (AMP)	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.55	Training for an authorized nuclear pharmacist (ANP)	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.57	Training for experienced RSO, teletherapy or medical physicist, AMP, authorized user (AU), nuclear pharmacist, and ANP	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.59	Recentness of training	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other	5	The guidance document provides additional information specific to Eye90 microspheres®.
Subpart C – General Technical Requirements					
35.60	Possession, use, and calibration of instruments used to measure the activity of unsealed byproduct material	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other	6.1	Although Eye90 microspheres® are not considered unsealed byproduct material, their pre-treatment activity is measured as per 35.60.
35.61	Calibration of survey instruments	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.63	Determination of dosages of unsealed byproduct material for medical use	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other	6.1	Although Eye90 microspheres® are not considered unsealed byproduct material, their pre-treatment activity is measured as per 35.60.

Appendix

Consolidated Technical Analysis

35.65	Authorization for calibration, transmission, and reference sources	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.67	Requirements for possession of sealed sources and brachytherapy sources	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.5	Microspheres are not managed as individual discrete sources and Y-90 is short-lived. 35.406, 20.1906, 20.1904, 35.69, 20.1801, and 35.92 are sufficient.
35.69	Labeling of vials and syringes	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.70	Surveys of ambient radiation exposure rate	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.8	The guidance document provides additional information specific to <i>Eye90 microspheres</i> [®] .
35.75	Release of individuals containing unsealed byproduct material or implants containing byproduct material	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.80	Provision of mobile medical service	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.92	Decay-in-storage	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	7.3	The guidance document provides additional information specific to <i>Eye90 microspheres</i> [®] , including concerns specific to isotopic impurities inherent in the production of Y-90.
Subpart F – Manual Brachytherapy					

Appendix
Consolidated Technical Analysis

35.400	Use of sources for manual brachytherapy	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.404	Surveys after source implant and removal	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.8	The guidance document provides additional information specific to Eye90 microspheres [®] . Eye90 microspheres [®] are permanent implants, therefore post-removal surveys are not applicable.
35.406	Brachytherapy sources accountability	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.5	The guidance document provides additional information specific to Eye90 microspheres [®] .
35.410	Safety instruction	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.415	Safety precautions	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.432	Calibration measurements of brachytherapy sources	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.1	Microspheres are too small to be calibrated in accordance with 35.432, therefore 35.63 and 35.60 are used for measuring each dosage.
35.433	Sr-90 sources for ophthalmic treatments	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.457	Therapy-related computer systems	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		

Appendix
Consolidated Technical Analysis

35.490	Training for use of manual brachytherapy sources	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	5	The guidance document provides additional information specific to <i>Eye90 microspheres</i> [®] .
35.491	Training for ophthalmic use of Sr-90	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
Subpart L – Records					
35.2024	Records of authority and responsibilities for radiation protection programs	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.2026	Records of radiation protection program changes	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.2040	Records of WDs	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.2041	Records for procedure for administrations requiring a WD	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.2060	Records of calibrations of instruments used to measure the activity of unsealed byproduct materials	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1566 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.2061	Records of radiation survey instrument calibrations	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®]		

Appendix

Consolidated Technical Analysis

The following table provides a list of 10 CFR Part 35 regulations and conditions the NRC has determined are applicable for use of *Eye90 microspheres*[®]. Licensees shall comply with all regulations which address use of *Eye90 microspheres*[®]. The table also provides specific conditions which the NRC has determined are necessary for the medical use of *Eye90 microspheres*[®]. In addition, the table lists where licensees and applicants can find additional guidance. Applicants may submit alternative list of regulations and specific conditions to be reviewed on a case-by-case basis by NRC staff.

Section	Description	Use Addressed in Regulation	Guidance	<i>Eye90 microspheres</i> [®] Guidance Section	Comment
35.1	Purpose and scope	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.2	Definitions	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.5	Maintenance of records	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.6	Provisions for the protection of human research subjects	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.7	FDA, other Federal, and State requirements	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		

Appendix

Consolidated Technical Analysis

Section	Description	Use Addressed in Regulation	Guidance	<i>Eye90 microspheres</i>[®] Guidance Section	Comment
35.8	Information collection requirements: OMB approval	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		

Appendix
Consolidated Technical Analysis

35.10	Implementation	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.11	License required	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.12	Application for license, amendment, or renewal	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.13	License amendments	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.14	Notifications	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.15	Exemptions regarding Type A specific licenses of broad scope	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.18	License issuance	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.19	Specific exemptions	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		

Appendix
Consolidated Technical Analysis

Subpart B – General Administrative Requirements					
35.24	Authority and responsibilities for the radiation protection program	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.26	Radiation protection program changes	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.27	Supervision	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.40	Written directives (WDs)	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.2	35.40(b)(6) details written directive requirements for permanent implant brachytherapy. 35.40(b)(6)(ii) requires a record of the number of sources implanted, which is not possible due to the small size and large number of microspheres involved in Eye90 therapy. Therefore there are different items required by the written directive.
35.41	Procedures for administrations requiring a WD	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.2	The guidance document provides additional information specific to Eye90 Microspheres.
35.49	Suppliers for sealed sources or devices for medical use	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.50	Training for Radiation Safety Officer (RSO) and Associate RSO	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®]	5.2	The guidance document provides additional information specific to Eye90 Microspheres.

Appendix
Consolidated Technical Analysis

			<input type="checkbox"/> Other		
35.51	Training for an authorized medical physicist (AMP)	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.55	Training for an authorized nuclear pharmacist (ANP)	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.57	Training for experienced RSO, teletherapy or medical physicist, AMP, authorized user (AU), nuclear pharmacist, and ANP	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.59	Recentness of training	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other	5	The guidance document provides additional information specific to Eye90 microspheres®.
Subpart C – General Technical Requirements					
35.60	Possession, use, and calibration of instruments used to measure the activity of unsealed byproduct material	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other	6.1	Although Eye90 microspheres® are not considered unsealed byproduct material, their pre-treatment activity is measured as per 35.60.
35.61	Calibration of survey instruments	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.63	Determination of dosages of unsealed byproduct material for medical use	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other	6.1	Although Eye90 microspheres® are not considered unsealed byproduct material, their pre-treatment activity is measured as per 35.60.

Appendix

Consolidated Technical Analysis

35.65	Authorization for calibration, transmission, and reference sources	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.67	Requirements for possession of sealed sources and brachytherapy sources	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.5	Microspheres are not managed as individual discrete sources and Y-90 is short-lived. 35.406, 20.1906, 20.1904, 35.69, 20.1801, and 35.92 are sufficient.
35.69	Labeling of vials and syringes	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.70	Surveys of ambient radiation exposure rate	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.8	The guidance document provides additional information specific to <i>Eye90 microspheres</i> [®] .
35.75	Release of individuals containing unsealed byproduct material or implants containing byproduct material	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.80	Provision of mobile medical service	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.92	Decay-in-storage	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	7.3	The guidance document provides additional information specific to <i>Eye90 microspheres</i> [®] , including concerns specific to isotopic impurities inherent in the production of Y-90.
Subpart F – Manual Brachytherapy					

Appendix
Consolidated Technical Analysis

35.400	Use of sources for manual brachytherapy	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.404	Surveys after source implant and removal	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.8	The guidance document provides additional information specific to Eye90 microspheres [®] . Eye90 microspheres [®] are permanent implants, therefore post-removal surveys are not applicable.
35.406	Brachytherapy sources accountability	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.5	The guidance document provides additional information specific to Eye90 microspheres [®] .
35.410	Safety instruction	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.415	Safety precautions	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.432	Calibration measurements of brachytherapy sources	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.1	Microspheres are too small to be calibrated in accordance with 35.432, therefore 35.63 and 35.60 are used for measuring each dosage.
35.433	Sr-90 sources for ophthalmic treatments	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.457	Therapy-related computer systems	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		

Appendix
Consolidated Technical Analysis

35.490	Training for use of manual brachytherapy sources	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	5	The guidance document provides additional information specific to <i>Eye90 microspheres</i> [®] .
35.491	Training for ophthalmic use of Sr-90	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
Subpart L – Records					
35.2024	Records of authority and responsibilities for radiation protection programs	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.2026	Records of radiation protection program changes	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.2040	Records of WDs	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.2041	Records for procedure for administrations requiring a WD	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.2060	Records of calibrations of instruments used to measure the activity of unsealed byproduct materials	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1566 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.2061	Records of radiation survey instrument calibrations	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®]		

Appendix
Consolidated Technical Analysis

			<input type="checkbox"/> Other		
35.2063	Records of dosages of unsealed byproduct material for medical use	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.2067	Records of leaks tests and inventory of sealed sources and brachytherapy sources	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other	4.4, 6.5	Microspheres are not managed as individual discrete sources and Y-90 is short-lived. 35.406, 20.1906, 20.1904, 35.69, 20.1801, and 35.92 are sufficient.
35.2070	Records of surveys for ambient radiation exposure rate	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.2075	Records of the release of individuals containing unsealed byproduct material or implants containing byproduct material	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.2080	Records of mobile medical services	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		
35.2092	Records of decay-in-storage	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other	7.3	The guidance document provides additional information specific to Eye90 microspheres®.
35.2204	Records of Mo-99, Sr-82, and Sr-85 concentrations	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> Eye90 microspheres® <input type="checkbox"/> Other		

Appendix

Consolidated Technical Analysis

35.2404	Records of surveys after source implant and removal	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.8	The guidance document provides additional information specific to Eye90 microspheres [®] . Eye90 microspheres [®] are permanent implants, therefore post-removal surveys are not applicable.
35.2406	Records of brachytherapy source accountability	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.5	Microspheres are not managed as individual discrete sources and Y-90 is short-lived. 35.406, 20.1906, 20.1904, 35.69, 20.1801, and 35.92 are sufficient.
35.2432	Records of calibration measurements of brachytherapy sources	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.1	Microspheres are too small to be calibrated in accordance with 35.432, therefore 35.63 and 35.60 are used for measuring each dosage.
35.2433	Records of decay of Sr-90 sources for ophthalmic treatments	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
Subpart M – Reports					
35.3045	Report and notification of a medical event	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input checked="" type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other	6.3	The guidance document provides additional information specific to Eye90 microspheres [®] .
35.3047	Report and notification of a dose to an embryo/fetus or a nursing child	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.3067	Report of a leaking source	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.3204	Report and notification for an eluate exceeding permissible Mo-99, Sr-	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	<input type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®]		

Appendix
Consolidated Technical Analysis

	82, and Sr-85 concentrations		<input type="checkbox"/> Other		
Subpart N – Enforcement					
35.4001	Violations	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
35.4002	Criminal penalties	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	<input checked="" type="checkbox"/> 1556 Vol 9 <input type="checkbox"/> <i>Eye90 microspheres</i> [®] <input type="checkbox"/> Other		
Additional Considerations					